

Reconfigurable Microfluidic System Architecture Based on Two-Dimensional Electrowetting Arrays

Jie Ding, Krishnendu Chakrabarty and Richard B. Fair
Department of Electrical and Computer Engineering
Duke University
{jding,krish,rfair}@ee.duke.edu

ABSTRACT

We present an architectural design and optimization methodology for performing biochemical reactions using two-dimensional electrowetting arrays. We define a set of basic microfluidic operations and leverage electronic design automation principles for system partitioning, resource allocation, and operation scheduling. Fluidic operations are carried out by properly configuring a set of grid points. While concurrency is desirable to minimize processing time, the size of the two-dimensional array limits the number of concurrent operations of any type. Furthermore, functional dependencies between the operations also limit concurrency. We use integer linear programming to minimize the processing time by automatically extracting parallelism from a biochemical assay. As a case study, we apply our optimization method to the polymerase chain reaction.

Keywords: Architectural optimization, integer linear programming, microelectrofluidics, partition map, reconfigurable architecture, scheduling, electrowetting.

1 INTRODUCTION

Electrowetting-based actuation for microelectrofluidics (MEFS) has recently been proposed for optical switching [1], chemical analysis [2], and rotating yaw rate sensing [3]. Pollack et al recently demonstrated that by varying the electrical potential along a linear array of electrodes, electrowetting techniques can be used to move liquid droplets along this line of electrodes [2]. By carefully controlling the electrical potential applied to the electrodes, fluid droplets can be moved as fast as 3 cm/sec.

Electrowetting can also be used to move droplets in a two-dimensional electrode array. By controlling the voltage on the electrodes, fluid droplets can be moved freely to any location on a two dimensional plane [2]. Fluid droplets can also be confined to a fixed location and isolated from other droplets moving around it.

Using two-dimensional electrowetting arrays, many useful microfluidic operations can be performed, such as storing, mixing and droplet splitting. The store operation is performed by applying a voltage to a isolated electrode to hold a droplet. This is analogous to a well. The voltage prevents this droplet from leaving the isolated electrode. The mix operation is performed by routing two droplets to the adjacent locations, where they are merged into one

droplet. Since the size of a droplet is kept small, effective mixing can be achieved by rapid fluidic diffusion during merging. Finally, the split operation is performed by creating opposite surface tension at the two ends of a fluid droplet and pulling it into two smaller droplets.

While two-dimensional electrowetting arrays are especially useful for biochemical analysis, system level design methodologies are required to harness this exciting new technology. In this paper, we leverage electronic design automations techniques to develop the first system-level design methodology for reconfigurable MEFS-based lab-on-a-chip.

Reconfigurable computing systems based on field-programmable gate-arrays (FPGAs) are now commonplace [4]. However, the “programmability” of FPGAs is limited by the well-defined roles of interconnect and logic blocks. Interconnect cannot be used for storing information, and logic blocks cannot be used for routing. In contrast, the MEFS architecture that we are developing offers significantly more programmability. The grid points in an array can be used for storage, functional operations, as well as for transporting fluid droplets. Therefore, partitioning, resource allocation, and scheduling have emerged as major challenges for system-level MEFS design targeted at a set of biochemical applications.

We have developed the syntax and semantics of microfluidic operations such as *MOVE*, *MIX*, and *SPLIT* that can be used to describe biochemical processes such as Polymerase Chain Reaction (PCR) [8]. The various fluid samples represent the operands. Such a *microfluidic program* must then be mapped to the two-dimensional array that represents the datapath of a *microfluidic computer*. (A separate electronic control unit drives the electrodes.) The execution of microfluidic operations requires the availability of datapath resources (set of grid points) that can be appropriately configured. For example, the *MIX* operation requires that a set of grid points be properly configured to act as a mixer. The size of the two-dimensional array limits the number of concurrent operations of any type that can be carried out. Furthermore, functional dependencies between the operations in a microfluidic program also limit concurrency.

The organization of the paper is as follows. In Section 2, we describe the two-dimensional electrowetting array and introduce the concepts of virtual microfluidic components and partition maps. Section 3 presents the scheduling problem for biochemical analysis and describes an integer linear programming approach for scheduling

Report Documentation Page				Form Approved OMB No. 0704-0188	
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.					
1. REPORT DATE 2005		2. REPORT TYPE		3. DATES COVERED -	
4. TITLE AND SUBTITLE Reconfigurable Microfluidic System Architecture Based on Two-Dimensional Electrowetting Arrays				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S)				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Defense Advanced Research projects Agency,3701 North Fairfax Drive,Arlington,VA,22203-1714				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT see report					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT	18. NUMBER OF PAGES 5	19a. NAME OF RESPONSIBLE PERSON
a. REPORT unclassified	b. ABSTRACT unclassified	c. THIS PAGE unclassified			

under resource constraints in two-dimensional electrowetting arrays. Finally, in Section 4 we present a case study based on the PCR reaction.

2 TWO-DIMENSIONAL ELECTROWETTING ARRAYS

A two-dimensional electrowetting array consists of a grid of electrodes on a two-dimensional plane (Figure 1). Fluid droplets are introduced to the device from the I/O ports on the boundary of the array. Droplets in the array have identical volumes. Hence, this type of device is also called a *unit-flow device*. It is desirable to maintain the unit-flow constraint since the rate of chemical and biomedical reactions grows exponentially with the growth of droplet volume [4].

Operations such as *STORE*, *MOVE*, *MIX*, and *SPLIT* are performed by controlling the electrical potential applied to the electrodes. It is easy to see that some of these operations violate the unit-flow assumption. For example, the fluid droplet size will double as a result of a mixing operation. Therefore, we always perform a split operation after mixing to maintain the droplet volume.

In a continuous-flow MEFS system, mixing is performed using a micromixer [5]. This is a specific device located at a fixed place in the microfluidic system. In unit-flow systems however, mixing operations can happen anywhere on the array, not necessarily at a specific location. If we define a mixer as the location where fluids mix, then a unit-flow mixer can be located at any arbitrary cell(s) in the electrode grid. This property is referred to as *reconfigurability*, and it is in many ways similar to the reconfigurability provided by FPGAs. However, as discussed in section 1, unit-flow devices allow a higher degree of reconfigurability than FPGAs. Storage cells, mixers and splitters can be created, removed, and relocated at runtime. This allows us to create extremely flexible and efficient biochemical analysis systems.

An abstract model of the unit flow system with a two-dimensional grid of electrodes is shown in Figure 1. A ground plane is positioned above the electrode array at a spacing that is less than the diameter of the droplets. I/O ports are placed at the boundary of the system.

2.1 Virtual devices and partition maps

In the unit-flow environment, the routes that droplets travel and the rendezvous points of fluid droplets are programmed into a micro-controller that controls the voltages of electrodes. The storage and interconnect on the datapath are viewed as virtual devices by the controller.

A virtual device is defined to have three regions (Figure 1). The first is the *functional region*, where a particular function is performed. The second type of region is called the *segregation region*, which wraps around the functional region. This insulates the functional region from its environment. The outer-most region of the device is the

inherited communication path. This provides a one-cell wide communication path for fluid droplet movement. Figure 1 shows a unit-flow storage cell. One droplet of a fluid sample is stored in each functional cell.

A *partition map* shows the time-varying positions of all the virtual devices inside the defined area. It is generated by the designer, and pre-loaded into the microcontroller, which then controls the electrode voltages according to partition map.

A partition map is similar to a virtual device, in that it is also a virtual map, and it only exists in the microcontroller specification. It is also dynamic in nature since it may change with time. Reconfiguration occurs when a new partition map is loaded into the controller. Figure 2 shows a partition map containing two storage cells, one input cell, and one mixer. (The labels A, B, ..., J, K will be explained later.) The inherited communication paths of adjacent devices are combined to form a single channel in the electrode array. This channel is used for fluid droplet transfer, and is called a communication path. It forms the main network for fluid movement. Researches have recently shown that it is possible to move the fluid droplets at a speed of 20 grids/second along this communication path [2]. The actual route along which a droplet moves is pre-determined and loaded into controller. If the routes of several consecutive droplets do not overlap, they are called *compatible routes*. Movements along compatible routes can be performed in parallel. If the routes are not compatible, the corresponding droplet movements must be performed sequentially.




Figure 1: A unit-flow storage device.

We define following operations that can be performed by virtual devices on a partition map.

- *MIX mixer_name*, where mixer_name is a reference to a particular mixer in the partition map.
- *SPLIT mix_name*, where mixer_name is a reference to a particular mixer in the partition map.
- *INPUT port_name, fluid_name*, where port_name is a reference to a port in the partition map.
- *MOVE source_name, destine_name, route_name*, where route_name is a reference to a pre-defined path.
- *PATH route_name, P1-P2-...-Pn*, defines a path for droplet movement

We next present a scheduling method for minimizing the processing time for fluid samples. We determine an

optimal sequence of fluidic operations to minimize completion time under resource constraints (availability of virtual devices) and dependencies between operations.




Figure 2: Partition map with two storage units, one input cell, and one mixer.

In contrast to droplet movement, fluidic operations such as *MIX* and *SPLIT* are slow processes. The mixing by diffusion at the nanometer level takes about 1 minute for completion. During the same time period, a droplet can move along 1800 grid points. Therefore, we ignore droplet movement time for operation scheduling.

In order to schedule microfluidic operations such as *MIX* and *SPLIT*, we divide the time span between two consecutive reconfigurations into equal length *time slots*. The length of a time slot equals the greatest common divisor of all the operations. For example, if a *MIX* operation takes 3 minutes and a *SPLIT* operation takes 2 minutes, then the time slot is set to 1 minute. In this case, the *MIX* operation will take 3 slots, and the *SPLIT* operation will take 2 slots. In this way, we digitize the continuous fluid operation and the controller starts or completes an operation at the end of each time slot.

3 SCHEDULE OPTIMIZATION USING INTEGER LINEAR PROGRAMMING

The order of execution of microfluidic operations must be determined after carefully considering the dependencies between the operations and the availability of resources. While dependencies are imposed by the biochemical application, the resource constraints are imposed by the size of the two-dimensional electrowetting array and the availability of virtual devices. In this section, we use the dataflow graph model of high-level synthesis [6] to represent the scheduling problem and solve it using integer linear programming (ILP). The motivation for using ILP lies in the fact that it is a well-understood optimization method and we can leverage a number of public domain solvers [7].

First, each step of a biochemical process is represented using either a single microfluidic operation or a series of basic microfluidic operations. Each such instance of an operation forms a node in the dataflow graph. A directed edge from node u to node v indicates a dependency between the operations corresponding to u and v , i.e. the operation corresponding to u must be carried out before the operation corresponding to v . The goal of the scheduling problem is to determine the start times (time slots) of each operation so that the total completion time is minimized.

Let $x_{i,j}$ be a binary variable defined as follows:

$$x_{i,j} = \begin{cases} 1, & \text{if operation } i \text{ starts at time slot } j \\ 0, & \text{otherwise} \end{cases}$$

where $1 \leq i \leq N$, the number of operations (nodes in the dataflow graph), and $1 \leq j \leq M$, the maximum possible index for a time slot. Note that M can be trivially obtained by adding up the number of time slots required for all the operations. Note also that since each operation is scheduled

$$\text{exactly once, } \sum_{j=1}^M x_{i,j} = 1, 1 \leq i \leq N.$$

The starting time S_i for operation i can now be expressed in terms of the set of variables $\{x_{i1}, x_{i2}, \dots, x_{im}\}$. Assuming that each time slot is of length 1 unit, we get

$$S_i = \sum_{j=1}^M jx_{ij}.$$

Each operation i has an associated execution time d_i . If there exists a dependency edge between operation i and operation j , then $S_j \geq S_i + d_i$. Such dependencies generally arise from the fluid samples that are used in each step of the biochemical reaction. These fluid samples are similar to variables in traditional architectural synthesis.

Finally, we add resource constraints to the ILP model. Let a_k be an upper bound on the number of operations of type k . We now have the following set of constraints for each k :

$$\sum_{i \in T(k)} \sum_{j=l-d_i+1}^l x_{ij} \leq a_k, 1 \leq l \leq M$$

The objective of this optimization problem is to minimize the completion time of the last operation, i.e.

$$\text{minimize } \max_i \left\{ \sum_{j=1}^M jx_{ij} + d_i \right\}. \text{ This can be linearized as:}$$

$$\text{minimize } C \text{ subject to } C \geq \sum_{j=1}^M (jx_{ij}) + d_i, 1 \leq i \leq N.$$

The ILP model can be easily solved using public-domain solvers. In our work, we used the *lpsolve* package from Eindhoven University of Technology in Netherlands [7].

4 PCR EXAMPLE

In this section, we present a case study for operation scheduling using the PCR reaction. The PCR reaction includes three basic steps. The first is the input section. In this part, a number of fluid samples are input into the system. Next, these samples are combined using a pre-determined set of *MIX* operations. Note that these are implemented by interleaving *MOV*, *MIX*, and *SPLIT* operations. Finally, the sample mixture is sent off-chip for a series of heating steps.

The input samples for PCR include *Tris-HCl* (pH 8.3), *KCl*, *gelatin*, *bovine serum albumin*, *beosynucleotide triphosphate*, *a primer*, *AmpliTaq DNA polymerase*, and λ DNA.

4.1 System configuration

The first example system we use is shown in Figure 2. The system can perform moving, mixing and splitting for the PCR reaction. It consists of 9-by-9 array of grid cells. A dedicated I/O port is located at the edge of the system. We assume that the mixing of two fluid droplets takes 2 minutes, while the input operation takes 0.5 minutes. Since the mix operation is always followed by a split operation, the latter is not explicitly considered here. Instead, we assume that the time for a split is included in the time for a mix operation. The speed of fluid movement is assumed to be 20 grid cells per minute.

The partition map for this example is also given by Figure 2. In addition to the partition map, the droplet route plan and schedule of operations (to be determined next) must be loaded into the controller.

4.2 Optimal Scheduling

We now describe how an optimal schedule can be derived to minimize the processing time. First, we represent the PCR reaction as a series of basic steps. This corresponds to a specification outlined by a lab technician, and serves as a user program. The user program can either be a sequential enumeration of steps, or it can contain a limited amount of hand-extracted concurrency. We then generate the dataflow graph based on the functional dependencies between the operations (Figure 3). An optimized PCR reaction for the datapath of Figure 1 and the dataflow graph of Figure 3 is given below:

Time (minutes)	Operations
	<i>Definition section</i>
	Path path1, C-E-F-G-D
	Path path2, C-E-F-H-D'
	Path path3, C-E-I-A
	Path path4, C-E-J-B
	Path path5, D'-F-H-K
	Path path6, A-G-F-D'
	Path path7, B-H-F-D'

0	Load partition map <i>INPUT</i> Tris-HCl
0.5	<i>MOVE</i> C, D, path1 <i>INPUT</i> KCl
1	<i>MOVE</i> C, D', path2 <i>INPUT</i> gelatin <i>MIX</i> D and D'
1.5	move C, A, path3 <i>INPUT</i> bovine serum albumin
2	<i>MOVE</i> C, B, path4
3	<i>MOVE</i> D', K, path5 <i>MOVE</i> A, D', path6 <i>INPUT</i> beosynucleotide triphosphate <i>MIX</i> D and D'
3.5	<i>MOVE</i> C, A, path3
5	<i>MOVE</i> move D', K, path5 <i>MOVE</i> A, D', path6 <i>INPUT</i> primer <i>MIX</i> D and D'
5.5	<i>MOVE</i> C, A, path3
7	<i>MOVE</i> D', K, path5 <i>MOVE</i> A, D', path6 <i>INPUT</i> AmpliTaq DNA polymerase <i>MIX</i> D and D'
7.5	<i>MOVE</i> C, A, path3
9	<i>MOVE</i> D', K, path5 <i>MOVE</i> A, D', path6 <i>INPUT</i> λ DNA <i>MIX</i> D and D'
9.5	<i>MOVE</i> Move C, A, path3
11	<i>MOVE</i> D', K, path5 <i>MOVE</i> A, D', path6 <i>MIX</i> D and D'
13	<i>MOVE</i> D', K, path5 <i>MOVE</i> B, D', path7 <i>MIX</i> D and D'
15	<i>MOVE</i> D', K, path5

Table 1: Optimized PCR reaction based on the datapath of Figure 1.




Figure 3: Dependency graph with input and mix operations.

The optimized PCR program of Table 1 was easy to derive since there is only one mixer in the system. The total processing time using this schedule is 15 minutes. We next

show how the processing time can be decreased further and an optimal schedule derived using ILP.

Consider the partition map shown in Figure 3 with four mixers. This allows greater parallelism and demonstrates the advantage is using ILP to minimize the processing time. The following discussion presents the ILP model for this example in more detail.

The PCR program contains a total of 15 *INPUT* and *MIX* operations. From Table 1, we note that an upper bound on the processing time is 15 minutes. Each time slot is of length 0.5 minutes (the assumed time for an *INPUT* operation), hence an upper bound on the number of time slots is 30. To build the ILP model for this partition map, we define a set of decision variables as discussed in Section 3. Thus our ILP model uses $x_{1,j} \dots x_{15,j}$ as the decision variables, where $j=1,2,\dots,30$. The start time of each operation can be expressed as follows:

$$S_1 = x_{1,2} + 2x_{1,3} + \dots + 29x_{1,30}$$

$$S_2 = x_{2,2} + 2x_{2,3} + \dots + 29x_{2,30}$$

...

$$S_{15} = x_{15,2} + 2x_{15,3} + \dots + 29x_{15,30}$$

The dependency between instructions can be denoted using the following set of inequalities:




Figure 4: Partition map with four mixers for PCR reaction.

$$S_9 > S_1, S_2$$

$$S_{10} > S_3, S_4$$

$$S_{11} > S_5, S_6$$

...

$$S_{15} > S_{13}, S_{14}$$

Finally, the resource constraints can be represented as:

$$x_{1,1} + x_{2,1} + \dots + x_{15,1} < 4$$

$$x_{2,2} + x_{2,2} + \dots + x_{15,2} < 4$$

...

$$x_{30,15} + x_{30,15} + \dots + x_{30,15} < 4$$

We solved this ILP model using *lpsolve*. It took 10 minutes of CPU time on a Sun Ultra Sparc with a 333 MHz processor and 128 MB of RAM. The optimum processing time is 10 minutes, 50% faster than the PCR program of Table 1.

5 CONCLUSIONS

We have presented a novel architectural design and optimization methodology for performing biochemical reactions using two-dimensional electrowetting arrays. We have defined a set of basic microfluidic operations and leveraged electronic design automation principles for system partitioning, resource allocation, and operation scheduling. While concurrency is desirable to minimize processing time, it is limited by the size of the two-dimensional array and functional dependencies between operations. We have used integer linear programming to minimize the processing time by automatically extracting parallelism from a biochemical assay. As a case study, we have applied our optimization method to the polymerase chain reaction.

Acknowledgement: This work was supported by DARPA under contract No. F30602-98-2-0140.

REFERENCES

- [1] J.L. Jackel, S. Hackwood, J. J. Veslka, and G. Beni. "Electrowetting Switch for Multimode Optical Fibers", *Applied Optics*, vol. 22, no.11, pp. 1765-1770, 1999
- [2] M. Pollack, R. B. Fair and A. Shenderov, "Electrowetting-based actuation of liquid droplets for microfluidic applications", *Applied Physics Letters*, vol. 77, no. 11, pp. 1725-1726, July 2000.
- [3] R.Yates, C. Williams, C. Shearwood and P.Mellor, "A Micromachined Rotating Yaw Rate Sensor", *Proc. Micromachined Devices and Components II, SPIE Meeting*, pp. 161-168, 1996.
- [4] S. M. Trimberger, ed., *Field-Programmable Gate Array Technology*, Kluwer Academic Publishers, Norwell, MA, 1994.
- [5] J. R. Welty, C. E. Wicks and R. E. Wilson, *Fundamentals of Momentum, Heat, and Mass Transfer*, John Wiley & Sons, Inc, New York 1983.
- [6] G. De Micheli, *Synthesis and Optimization of Digital Circuits*, McGraw-Hill, Inc., New York, NY, 1994.
- [7] M. Berkelaar, *lpsolve 3.0*, Eindhoven University of Technology, Eindhoven, The Netherlands, ftp://ftp.ics.ele.tue.nl/pub/lp_solve.
- [8] L. C. Waters et al., "Multiple Sample PCR Amplification and Electrophoretic Analysis on a Microchip", *Analytical Chemistry*, vol. 70, no. 24, December 1998.